

WARNING

Potential for Incorrect Blood Glucose Reading

Patient Name

ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitor or test strips that utilize a glucose dehydrogenase flavin-dinnucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (Icodextrin 7.5%) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hypoglycemia, leading to life-threatening events. Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list.

is using **EXTRANEAL** (Icodextrin 7.5%)
peritoneal dialysis solution.

Baxter

PSTGL03IR3 Date of Preparation Nov 2019. ADV06/629R.

BAXTER and EXTRANEAL are trademarks of Baxter International Inc.

Baxter Healthcare Philippines Inc.

Silangan Industrial Estate Brgy. Canlubang Calamba City Laguna Philippines 4028
Tel +63 2 7238 8080 Fax +63 2 7238 8034