

See reverse side for warnings

PATIENT NAME

is using **Peritoneal Dialysis Solution**
with 7.5% Icodextrin (EXTRANEAL)

Emergency Contact Information

Nephrologist

()

Dialysis Unit/Center

()

Other Contact

()

BAXTER and EXTRANEAL are trademarks of Baxter International Inc. PH-002-2010

WARNING

Potential for Incorrect Blood Glucose Reading

ONLY use glucose monitors and test strips that are glucose specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxydoreductase methods. In addition, some but not all monitors or test strips that utilize a glucose dehydrogenase flavin adenine dinucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **Peritoneal Dialysis Solution with 7.5% Icodextrin (EXTRANEAL)** due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hyperglycemia, leading to life-threatening events.

Visit www.glucosafety.com for additional information, including a glucose monitor compatibility list. Call (02) 687-7772.

PH001-2010

Baxter

BAXTER and EXTRANEAL are trademarks of Baxter International Inc.