

See reverse side for warnings

Patient Name

is using **EXTRANEAL** (Icodextrin 7.5%)  
peritoneal dialysis solution.

**Emergency Contact Information**

**Dialysis Unit** \_\_\_\_\_

**Other Contact** \_\_\_\_\_

**BAXTER and EXTRANEAL are trademarks of Baxter International Inc.  
PSTGL04IR3. Date of Preparation November 2019. ADV06/628R.**

**Baxter Healthcare Philippines Inc.**

Silangan Industrial Estate Brgy. Canlubang Calamba City Laguna Philippines 4028  
Tel +63 2 7238 8080 Fax +63 2 7238 8034

SG/81/19-0002

# WARNING :

## Potential for Incorrect Blood Glucose Reading

**ONLY use glucose monitors and test strips that are glucose-specific. These methods are common in clinical laboratories. Contact the manufacturer of the glucose monitors and test strips to determine the method that is used.**

DO NOT use glucose monitors or test strips that utilize glucose dehydrogenase pyrroloquinolinequinone (GDH-PQQ) or glucose-dye-oxidoreductase methods. In addition, some but not all monitor or test strips that utilize a glucose dehydrogenase flavin-dinnucleotide (GDH-FAD) method should not be used. Use of these methods may result in falsely elevated blood glucose readings in patients using **EXTRANEAL** (Icodextrin 7.5%) due to maltose interference. Falsely elevated blood glucose readings may mask true hypoglycemia or lead to the erroneous diagnosis of hypoglycemia, leading to life-threatening events.

Visit [www.glucosesafety.com](http://www.glucosesafety.com) for additional information, including a glucose monitor compatibility list.

***Baxter***